

**IK5014 Human ADAMTS13 ELISA Kit**


Size: 96wells/kit, with removable strips.

**Sandwich High Sensitivity ELISA kit for Quantitative Detection of Human ADAMTS13. 96wells/kit, with removable strips.**Typical Data Obtained from Human ADAMTS13  
(TMB reaction incubate at 37°C for 25-30min)

| Concentration(ng/ml) | 0     | 0.78  | 1.56  | 3.12  | 6.25  | 12.5  | 25    | 50    |
|----------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| O.D.                 | 0.051 | 0.101 | 0.134 | 0.243 | 0.439 | 0.829 | 1.482 | 2.472 |

**Typical Human ADAMTS13 ELISA Kit Standard Curve**

This standard curve was generated at IS for demonstration purpose only. A standard curve must be run with each assay.

**Range:** 0.78ng/ml-50ng/ml**Sensitivity:** <20pg/ml**Specificity:** Natural and recombinant Human ADAMTS13**Cross-reactivity:** There is no detectable cross-reactivity with other relevant proteins.**Storage:** Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles  
(Shipped with wet ice.)**Intra/Inter Assay Precision:****Intra-Assay Precision** (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision.**For Research use only**  
**IMMUNOLOGICAL SCIENCES**Web-site: [www.immunologicalsciences.com](http://www.immunologicalsciences.com) - E-Mail: [info@immunologicalsciences.com](mailto:info@immunologicalsciences.com)

**Inter-Assay Precision** (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision.

| Sample             | Intra-Assay Precision |       |      | Inter-Assay Precision |       |      |
|--------------------|-----------------------|-------|------|-----------------------|-------|------|
|                    | 1                     | 2     | 3    | 1                     | 2     | 3    |
| n                  | 16                    | 16    | 16   | 24                    | 24    | 24   |
| Mean(ng/ml)        | 5.62                  | 10.1  | 25.5 | 6.23                  | 12.1  | 26.4 |
| Standard deviation | 0.259                 | 0.444 | 1.07 | 0.467                 | 0.835 | 1.95 |
| CV(%)              | 4.6                   | 4.4   | 4.2  | 7.5                   | 6.9   | 7.4  |

## Assay Principle

IS' Human ADAMTS13 ELISA Kit was based on standard sandwich enzyme-linked immunosorbent assay technology. A monoclonal antibody from mouse specific for ADAMTS13 has been precoated onto 96-well plates. Standards(CHO, Q34-W688) and test samples are added to the wells, a biotinylated detection polyclonal antibody from goat specific for ADAMTS13 is added subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human ADAMTS13 amount of sample captured in plate.

## Kit Components

### Materials included in the kit

| Description   | Quantity              |
|---|-----------------------|
| 96-well plate precoated with anti-Human ADAMTS13 antibody | 1                     |
| lyophilized recombinant Human ADAMTS13 standard           | 50ng/tubex2           |
| biotinylated anti-Human ADAMTS13 antibody                 | 130ul(dilution 1:100) |
| Avidin-Biotin-Peroxidase Complex(ABC)                     | 130ul(dilution 1:100) |
| sample diluent buffer                                     | 30ml                  |
| antibody diluent buffer                                   | 12ml                  |
| ABC diluent buffer  | 12ml                  |
| TMB color developing agent                                | 10ml                  |
| TMB stop solution   | 10ml                  |
| Adhesive cover  | 4                     |
| PBS washing buffer  | Powder for 1000ml     |

Washing buffer Preparation: Dissolve PBS Wash B to 1000ml distilled water and adjust pH to 7.2~7.6. Finally, adjust the total volume to 1L.

**For Research use only**  
**IMMUNOLOGICAL SCIENCES**

Web-site: [www.immunologicalsciences.com](http://www.immunologicalsciences.com) - E-Mail: [info@immunologicalsciences.com](mailto:info@immunologicalsciences.com)

**Materials Required But Not Provided**

1. Microplate reader in standard size.
2. Automated plate washer.
3. Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection.
4. Clean tubes and Eppendorf tubes.

**Notice Before Application**

Please read the following instructions before starting the experiment.

1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
2. The TMB Color Developing agent is colorless and transparent before using, contact us freely if it is not the case.
3. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
4. Duplicate well assay is recommended for both standard and sample testing.
5. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
6. Don't reuse tips and tubes to avoid cross contamination.
7. Avoid using the reagents from different batches together.
8. In order to avoid marginal effect of plate incubation due to temperature difference (reaction may be stronger in the marginal wells), it is suggested that the diluted ABC and TMB solution will be pre-warmed in 37°C for 30 min before using.
9. Take precautionary measures to prevent operator contamination (such as saliva and other body fluids) of kit reagents while running this assay.

**Preparation****1. Sample Preparation and Storage**

Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

- **Serum:** Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store samples at -20°C.
- **Cell culture supernates:** Remove particulates by centrifugation, assay immediately or aliquot and store samples at -20°C.
- **Plasma:** Collect plasma using heparin, EDTA or citrate as an anticoagulant. Centrifuge for 15 min at 1500 x g within 30 min of collection. Assay immediately or aliquot and store samples at -20°C.

**Sample Dilution Guideline**

The user needs to estimate the concentration of the target protein in the sample and select a proper dilution factor so that the diluted target protein concentration falls near the middle of the linear regime in the standard curve. Dilute the sample using the provided diluent buffer. The following is a guideline for sample dilution. Several trials may be necessary in practice. **The sample must be well mixed with the diluents buffer.**

**For Research use only**  
**IMMUNOLOGICAL SCIENCES**

Web-site: [www.immunologicalsciences.com](http://www.immunologicalsciences.com) - E-Mail: [info@immunologicalsciences.com](mailto:info@immunologicalsciences.com)



- **High target protein concentration (500ng/ml-5000ng/ml).** The working dilution is 1:100. i.e. Add 1µl sample into 99 µl sample diluent buffer.
- **Medium target protein concentration (50ng/ml-500ng/ml).** The working dilution is 1:10. i.e. Add 10µl sample into 90 µl sample diluent buffer.
- **Low target protein concentration (0.78ng/ml-50ng/ml).** The working dilution is 1:2. i.e. Add 50µl sample to 50 µl sample diluent buffer.
- **Very low target protein concentration (0ng/ml-0.78ng/ml).** No dilution necessary, or the working dilution is 1:2.

### 3. Reagent Preparation and Storage

- A. Reconstitution of the Human ADAMTS13 standard: ADAMTS13 standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of ADAMTS13 standard (50ng/tube) are included in each kit. Use one tube for each experiment.
- a. 50ng/ml of Human ADAMTS13 standard solution: Add 1ml sample diluent buffer into one tube, keep the tube at room temperature for 10 min and mix thoroughly.
  - b. 25ng/ml→0.78125ng/ml of Human ADAMTS13 standard solutions: Label 6 Eppendorf tubes with 25ng/ml, 12.5ng/ml, 6.25ng/ml, 3.125ng/ml, 1.5625ng/ml, 0.78125ng/ml respectively. Aliquot 0.3ml of the sample diluent buffer into each tube. Add 0.3ml of the above 50ng/ml ADAMTS13 standard solution into 1st tube and mix. Transfer 0.3ml from 1st tube to 2nd tube and mix. Transfer 0.3ml from 2nd tube to 3rd tube and mix, and so on.

**Note:** The standard solutions are best used within 2 hours. The 50ng/ml standard solution should be stored at 4°C for up to 12 hours, or at -20°C for up to 48 hours. Avoid repeated freeze-thaw cycles.

- B. Preparation of biotinylated anti-Human ADAMTS13 antibody working solution: The solution should be prepared no more than 2 hours prior to the experiment.
- a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
  - b. Biotinylated anti-Human ADAMTS13 antibody should be diluted in 1:100 with the antibody diluent buffer and mixed thoroughly. (i.e. Add 1µl Biotinylated anti-Human ADAMTS13 antibody to 99µl antibody diluent buffer.)
- C. Preparation of Avidin-Biotin-Peroxidase Complex (ABC) working solution: The solution should be prepared no more than 1 hour prior to the experiment.
- a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
  - b. Avidin- Biotin-Peroxidase Complex (ABC) should be diluted in 1:100 with the ABC dilution buffer and mixed thoroughly. (i.e. Add 1µl ABC to 99µl ABC diluent buffer.)

### Assay Procedure

The ABC working solution TMB color developing agent and TMB stop solution must be kept warm at 37°C for 30 min before use. When diluting samples and reagents, they must be mixed completely and evenly. Standard ADAMTS13 detection curve should be prepared for each experiment. The user will decide sample dilution fold by crude estimation of ADAMTS13 amount in samples.

1. Aliquot 0.1ml per well of the 50ng/ml, 25ng/ml, 12.5ng/ml, 6.25ng/ml, 3.125ng/ml, 1.5625ng/ml, 0.78125ng/ml Human ADAMTS13 standard solutions into the precoated 96-well plate. Add 0.1ml of the sample diluent buffer into the control well (Zero well). Add 0.1ml of each properly diluted sample of Human cell culture supernates, serum or plasma (heparin, EDTA, citrate) to each empty well. **See**

**For Research use only**  
**IMMUNOLOGICAL SCIENCES**

Web-site: [www.immunologicalsciences.com](http://www.immunologicalsciences.com) - E-Mail: [info@immunologicalsciences.com](mailto:info@immunologicalsciences.com)



**"Sample Dilution Guideline" above for details.** It is recommended that each Human ADAMTS13 standard solution and each sample be measured in duplicate.

2. Seal the plate with a new adhesive cover provided and incubate at 37°C for 90 min.
3. Remove the cover, discard plate content, and blot the plate onto paper towels or other absorbent material. Do NOT let the wells completely dry at any time.
4. Add 0.1ml of biotinylated anti-Human ADAMTS13 antibody working solution into each well, seal the plate with a new adhesive cover provided and incubate at 37°C for 60 min.
5. Wash plate 3 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (**Plate Washing Method:** Discard the solution in the plate without touching the side walls. Blot the plate onto paper towels or other absorbent material. Soak each well with at least 0.3 ml PBS or TBS buffer for 1~2 minutes. Repeat this process two additional times for a total of THREE washes. Note: For automated washing, aspirate all wells and wash THREE times with PBS or TBS buffer, overfilling wells with PBS or TBS buffer. Blot the plate onto paper towels or other absorbent material.)
6. Add 0.1ml of prepared ABC working solution into each well, seal the plate with a new adhesive cover provided and incubate at 37°C for 30 min.
7. Wash plate 5 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1-2 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (See Step 5 for plate washing method.)
8. Add 90µl of prepared TMB color developing agent into each well, seal the plate with a new adhesive cover and incubate at 37°C in dark for 25-30min (**Note:** For reference only, the optimal incubation time should be determined by end user. And the shades of blue can be seen in the wells with the four most concentrated Human ADAMTS13 standard solutions; the other wells show no obvious color).
9. Add 0.1ml of prepared TMB stop solution into each well. The color changes into yellow immediately.
10. Read the O.D. absorbance at 450nm in a microplate reader within 30 min after adding the stop solution.

For calculation, (the relative O.D.450) = (the O.D.450 of each well) – (the O.D.450 of Zero well). The standard curve can be plotted as the relative O.D.450 of each standard solution (Y) vs. the respective concentration of the standard solution (X). The Human ADAMTS13 concentration of the samples can be interpolated from the standard curve.

Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.

#### **Summary**

1. Add samples and standards and incubate the plate at 37°C for 90 min. Do not wash.
2. Add biotinylated antibodies and incubate the plate at 37°C for 60 min. Wash plate 3 times with 0.01M TBS.
3. Add ABC working solution and incubate the plate at 37°C for 30 min. Wash plate 5 times with 0.01M TBS.
4. Add TMB color developing agent and incubate the plate at 37°C in dark for 25-30min.
5. Add TMB stop solution and read.

#### **Background**

ADAMTS13 is a zinc-containing metalloprotease enzyme. The ADAMTS13 gene is mapped to chromosome 9q34 by genomic sequence analysis. It is reported that a metal-containing proteolytic enzyme (metalloprotease) in normal plasma cleaves the peptide bond between tyrosine at position 842 and methionine at position 843 in monomeric subunits of von Willebrand factor, thereby degrading the large multimers. And it is confirmed that the ADAMTS13 gene encodes the von Willebrand factor-cleaving protease (VWFCP). ADAMTS13 is secreted in blood and degrades large vWf multimers, decreasing their activity.

**For Research use only  
IMMUNOLOGICAL SCIENCES**

Web-site: [www.immunologicalsciences.com](http://www.immunologicalsciences.com) - E-Mail: [info@immunologicalsciences.com](mailto:info@immunologicalsciences.com)